



P. P. H. U. BEROTEX Jaroslaw Majtczak
ul. Zygmunta 56
93-331 Lodz
Poland

Your notice of
 24-06-2020

Your reference

Date
 01-07-2020

Analysis Report 20.03950.02


Required tests :

EN 14126 (2003) + AC

ISO 16603 (2004)

Resistance to penetration by blood and body fluids

Identification number	Information given by the client	Date of receipt
T2013842	BERMED LAMINATE 53 GSM MEDICAL GREEN	24-06-2020



Sylvie Niessen
 Order responsible

This report may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel.
 The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.
 In assessing compliance with the specifications, we did not take into account the uncertainty on the test results.






Reference: T2013842 - BERMED LAMINATE 53 GSM MEDICAL GREEN

Resistance to penetration by blood and body fluids

Date of ending the test	26-06-2020
Standard used	ISO 16603 (2004)
Product standard	EN 14126 (2003) + AC (2004)
Type of sample	Laminated green non woven
Dimension of the test specimens	7.5cm x 7.5cm
Number of test specimens	3
Paraffin-sealed edges	No
Sampling	Everywhere in the received piece of non woven
Test specimens conditioning	21 ± 5°C and 60 ± 10%RH The sample is not tested in conditioned area but directly after conditioning.
Sterilization	None
Pre treatment performed	None
Body fluid used	Synthetic blood
Side in contact with the synthetic blood	Laminated side
Test procedure used	Procedure D (5 min 0kPa + 5 min 1.75 kPa + 5 min 3.5 kPa + 5 min 7 kPa + 5 min 14kPa + 5 min 20kPa - With screen)
Retaining screen specifications	Metal square mesh screen (open area >50%), limiting the deflection of the sample to ≤ 5.0 mm
Method used to improve the visualisation	None
Surface tension of synthetic blood	0.042 ± 0.002 N/m
Temperature during the test	22°C



Results

Pass / Fail results

Test specimen	0 kPa	1.75 kPa	3.5 kPa	7 kPa	14 kPa	20 kPa
#1	Pass	Pass	Pass	Pass	Pass	Pass
#2	Pass	Pass	Pass	Pass	Pass	Pass
#3	Pass	Pass	Pass	Pass	Pass	Pass

Pass = no penetration of synthetic blood

Fail = synthetic blood penetration



bermedo non woven